



Gaseous detector device for an environmental electron probe microanalyzer (G D Danilatos, CSIRO Division of Textile Physics)

Author: G D Danilatos, CSIRO Division of Textile Physics, 338

Research Disclosure Database Number 233011
(Research Disclosure Journal Number : 23311)

Published in September 1983

The Research Disclosure Journal is normally published and distributed on the 10th of every month unless that date coincides with a weekend or public holiday, when it is published directly afterwards. In these cases it is always published by the 12th of every month. Every disclosure is also placed on the RD Electronic database as soon as it is received and it may be published on the database prior to being published in the next edition of Journal.

Research Disclosure is the unique international defensive publication service that allows the world's intellectual property community to establish prior art, and provides an alternative to obtaining a patent at a fraction of the cost and the time taken. It is the world's longest running, independent, industry standard prior art disclosure service.

Kenneth Mason Publications Ltd give consent for this disclosure to be printed out providing it is for personal use, or for the personal or internal use of patent examiners or specific clients only. Photocopies may be made providing it is for personal use, or for the personal or internal use of patent examiners or specific clients and not for resale and the copier pays the usual photocopying fee/s to the relevant Copyright Clearance Centre. This consent does not extend to abstracting for general distribution for advertising, or promotional purposes, for creating new collective works or for resale. This consent also does not extend to other kinds of scanning, printing or copying, such as printing, scanning or copying for general distribution for advertising, or promotional purposes, for creating new collective works or for resale. Document delivery services are expressly forbidden from scanning, printing or copying any Research Disclosure content for re-sale unless specifically licensed to do so by the publishers.

Research Disclosure Journal, ISSN 0374-4353. © Kenneth Mason Publications Ltd
The Book Barn, Westbourne, Hants. PO10 8RS. UK
Tel: +44 (0)1243-377977 Fax: +44 (0)1243-379136
e-mail : info@ResearchDisclosure.com

ABSTRACT

The electron beam in instruments such as electron microscopes or general electron probe micro-analyzers (EPMA) interacts with the specimen generating as a result a variety of signals such as secondary electrons, backscattered electrons, transmitted electrons, X-rays, cathodoluminescence or absorbed current. Each of these signals is detected by appropriate detection means (detectors).

In conventional electron microscopy, both the electron beam and the specimen under examination are under high vacuum. However, it is possible to examine the specimen inside a gaseous environment if special known techniques¹ are employed to separate the high pressure region of the specimen from the high vacuum of the electron optics column.

It has now been found that the gas itself in an environmental EPMA can be used for the detection of signals. Signals capable of ionizing the gas create positive and negative charges which can be collected and monitored by suitable means. For example, a wire inserted in the gaseous environment and biased with a positive or negative voltage can collect the negative or positive charges (ions) and transmit a current into a video display, corresponding to the properties of the area of interaction between the electron beam and the specimen. Images of specimens have been produced in the environmental scanning electron microscope by using the ions produced by the ionizing signals. These images can correspond to the backscattered electron images at a certain gas pressure, but additional information appears as the gas pressure varies. The information on the images varies as different signals are detected with variation in the electrode configuration, bias, gas constitution and gas pressure.

In addition, the presence of gaseous and liquid phases around the specimen has made possible the recording of data from both insulators and conductors using the absorbed current mode of detection.

The main advantage of the gaseous detector device over the known detectors is that the gas can serve as a conditioning medium at the same time. Thus the specimens can be maintained at their natural state, they can be liquids, or they can take part in a physical or chemical reaction during their examination under the electron beam.

Again, various signals can be detected simultaneously. To achieve a similar result with conventional detectors, it is necessary to adapt separately each of these detectors for operation in the presence of gas. Such adaptation, however, is not always possible or practical. A further advantage is the extreme simplicity of the device.

REFERENCE

¹ Danilatos GD, Design and Construction of an Atmospheric or Environmental SEM, (Part 1), Scanning, 4, 9-20 (1981).

Disclosed by G D Danilatos, CSIRO Division of Textile Physics, 338 Blaxland Road, Ryde, NSW 2112, Australia